

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN101218\
 Data File : VN051853.D
 Acq On : 13 Oct 2018 00:27
 Operator : MD\SY
 Sample : PB113779ZHE#06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 34 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB113779ZHE#06

Manual Integrations
 APPROVED

MMDadoda
 10/15/2018 11:28:06 AM

Quant Time: Oct 13 03:42:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101118W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 09:51:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	806619	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1180840	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1038415	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	378340	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	435058	56.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.86%	
35) Dibromofluoromethane	7.59	113	432593	58.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.84%	
50) Toluene-d8	10.09	98	1612438	58.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.86%	
62) 4-Bromofluorobenzene	12.40	95	468935	49.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.30%	
Target Compounds						
16) Acetone	3.82	43	17512	10.53	ug/l	95
18) Methyl Acetate	4.34	43	11231	1.52	ug/l	99
20) Methylene Chloride	4.56	84	11436m	1.66	ug/l	
43) Isopropyl Acetate	8.17	43	267814	27.09	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN101218\
Data File : VN051853.D
Acq On : 13 Oct 2018 00:27
Operator : MD\SY
Sample : PB113779ZHE#06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 34 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB113779ZHE#06

Manual Integrations
APPROVED

MMDadoda
10/15/2018 11:28:06 AM

Quant Time: Oct 13 03:42:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101118W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 09:51:20 2018
Response via : Initial Calibration

