

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN101218\
 Data File : VN051854.D
 Acq On : 13 Oct 2018 00:51
 Operator : MD\SY
 Sample : PB113779ZHE#07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 35 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB113779ZHE#07

Manual Integrations
 APPROVED

MMDadoda
 10/15/2018 11:28:06 AM

Quant Time: Oct 13 03:44:20 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101118W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 09:51:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	857455	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1249281	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1098866	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	406477	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	472134	57.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.22%	
35) Dibromofluoromethane	7.59	113	457881	58.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.88%	
50) Toluene-d8	10.09	98	1723649	59.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.10%	
62) 4-Bromofluorobenzene	12.40	95	495319	49.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.14%	
Target Compounds						
16) Acetone	3.82	43	17076m	9.66	ug/l	Qvalue
18) Methyl Acetate	4.33	43	10690	1.25	ug/l #	87
20) Methylene Chloride	4.55	84	11804	1.61	ug/l #	85
43) Isopropyl Acetate	8.17	43	268006	25.60	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN101218\
Data File : VN051854.D
Acq On : 13 Oct 2018 00:51
Operator : MD\SY
Sample : PB113779ZHE#07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 35 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB113779ZHE#07

Manual Integrations
APPROVED

MMDadoda
10/15/2018 11:28:06 AM

Quant Time: Oct 13 03:44:20 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101118W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 09:51:20 2018
Response via : Initial Calibration

