

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN101218\
 Data File : VN051859.D
 Acq On : 13 Oct 2018 2:54
 Operator : MD\SY
 Sample : J5199-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 40 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 OK-02-101118

Quant Time: Oct 13 04:03:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101118W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 09:51:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	825716	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1193620	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1053214	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	413436	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	454991	57.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.30%	
35) Dibromofluoromethane	7.59	113	438982	59.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.30%	
50) Toluene-d8	10.09	98	1649349	59.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.28%	
62) 4-Bromofluorobenzene	12.40	95	484539	50.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.50%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	14507	8.52	ug/l	96
18) Methyl Acetate	4.34	43	13279	1.93	ug/l #	88
20) Methylene Chloride	4.55	84	99002	14.05	ug/l	90
40) Benzene	8.04	78	60974	2.19	ug/l	100
43) Isopropyl Acetate	8.17	43	282799	28.33	ug/l	94
52) Toluene	10.16	92	57578	3.08	ug/l	99
68) m/p-Xylenes	11.62	106	17105	1.20	ug/l	98
95) Naphthalene	15.13	128	181799	15.28	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN101218\
Data File : VN051859.D
Acq On : 13 Oct 2018 2:54
Operator : MD\SY
Sample : J5199-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 40 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
OK-02-101118

Quant Time: Oct 13 04:03:21 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101118W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 09:51:20 2018
Response via : Initial Calibration

