

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN101218\
 Data File : VN051860.D
 Acq On : 13 Oct 2018 3:19
 Operator : MD\SY
 Sample : J5193-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 41 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 BU-01-101118

Quant Time: Oct 13 04:05:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101118W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 09:51:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	815195	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1176935	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1028256	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	368126	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	450830	57.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.72%	
35) Dibromofluoromethane	7.59	113	437943	59.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.68%	
50) Toluene-d8	10.09	98	1616970	59.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.58%	
62) 4-Bromofluorobenzene	12.40	95	455396	48.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.74%	
Target Compounds						
16) Acetone	3.82	43	13838	8.23	ug/l #	86
18) Methyl Acetate	4.33	43	13500	2.02	ug/l #	80
20) Methylene Chloride	4.55	84	34840	5.01	ug/l	85
43) Isopropyl Acetate	8.17	43	199246	20.08	ug/l	96
95) Naphthalene	15.13	128	24018	6.01	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN101218\
Data File : VN051860.D
Acq On : 13 Oct 2018 3:19
Operator : MD\SY
Sample : J5193-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 41 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
BU-01-101118

Quant Time: Oct 13 04:05:30 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101118W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 09:51:20 2018
Response via : Initial Calibration

