

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101219\
 Data File : VN058688.D
 Acq On : 11 Oct 2019 9:17
 Operator : JC/SP
 Sample : VN1012WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1012WBL01

Quant Time: Oct 11 17:33:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N101119W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Oct 11 04:26:47 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.17	128	69868	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.57	114	380958	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	338696	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.01	65	186407	31.43	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	104.77%
60) 4-Bromofluorobenzene	12.40	95	161973	27.34	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	91.13%
63) Toluene-d8	10.08	98	482378	30.16	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	100.53%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101219\
Data File : VN058688.D
Acq On : 11 Oct 2019 9:17
Operator : JC/SP
Sample : VN1012WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1012WBL01

Quant Time: Oct 11 17:33:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N101119W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Oct 11 04:26:47 2019
Response via : Initial Calibration

