

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN101220\
 Data File : VN064078.D
 Acq On : 12 Oct 2020 15:23
 Operator : JC/MD
 Sample : VN1012MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1012MBL01

Quant Time: Oct 13 01:38:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	236227	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	427940	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	432125	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	158229	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.98	65	197686	50.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.30%	
35) Dibromofluoromethane	7.54	113	147577	49.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.40%	
50) Toluene-d8	10.06	98	553074	49.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.26%	
62) 4-Bromofluorobenzene	12.38	95	198423	46.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.78%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101220\
Data File : VN064078.D
Acq On : 12 Oct 2020 15:23
Operator : JC/MD
Sample : VN1012MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1012MBL01

Quant Time: Oct 13 01:38:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

