

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101220\
 Data File : VN064083.D
 Acq On : 12 Oct 2020 17:26
 Operator : JC/MD
 Sample : L4281-03ME
 Misc : 5.08g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 49-93ME

Quant Time: Oct 13 02:06:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	213629	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	393737	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	400450	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	181995	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.98	65	181431	51.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.82%	
35) Dibromofluoromethane	7.55	113	127307	46.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.20%	
50) Toluene-d8	10.06	98	495678	47.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.72%	
62) 4-Bromofluorobenzene	12.38	95	200426	51.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.96%	
Target Compounds						
16) Acetone	3.79	43	102477	94.324	ug/l	100
52) Toluene	10.13	92	38694	5.491	ug/l	99
68) m/p-Xylenes	11.59	106	15186	2.726	ug/l	99
69) o-Xylene	11.92	106	6826	1.311	ug/l	95
80) 1,3,5-Trimethylbenzene	12.71	105	28687	2.576	ug/l	98
84) 1,2,4-Trimethylbenzene	13.02	105	83951	7.550	ug/l	96
95) Naphthalene	15.11	128	9749	2.879	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101220\
Data File : VN064083.D
Acq On : 12 Oct 2020 17:26
Operator : JC/MD
Sample : L4281-03ME
Misc : 5.08g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
49-93ME

Quant Time: Oct 13 02:06:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

