

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101220\  
 Data File : VN064087.D  
 Acq On : 12 Oct 2020 19:02  
 Operator : JC/MD  
 Sample : L4225-09  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 SU-01-100920

Quant Time: Oct 13 02:32:05 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N093020W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Oct 01 01:46:38 2020  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc    | Units   | Dev(Min) |
|-----------------------------|-------|------|----------|---------|---------|----------|
| 1) Pentafluorobenzene       | 7.62  | 168  | 224555   | 50.00   | ug/l    | 0.00     |
| 34) 1,4-Difluorobenzene     | 8.55  | 114  | 417956   | 50.00   | ug/l    | 0.00     |
| 63) Chlorobenzene-d5        | 11.38 | 117  | 447502   | 50.00   | ug/l    | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.32 | 152  | 182509   | 50.00   | ug/l    | 0.00     |
| System Monitoring Compounds |       |      |          |         |         |          |
| 33) 1,2-Dichloroethane-d4   | 7.99  | 65   | 188307   | 50.76   | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =       | 101.52% |          |
| 35) Dibromofluoromethane    | 7.55  | 113  | 138154   | 47.64   | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =       | 95.28%  |          |
| 50) Toluene-d8              | 10.06 | 98   | 547646   | 49.81   | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =       | 99.62%  |          |
| 62) 4-Bromofluorobenzene    | 12.38 | 95   | 223786   | 54.15   | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =       | 108.30% |          |
| Target Compounds            |       |      |          |         |         |          |
| 16) Acetone                 | 3.78  | 43   | 168099   | 147.197 | ug/l    | 100      |
| 18) Methyl Acetate          | 4.28  | 43   | 5483     | 1.849   | ug/l #  | 77       |
| 20) Methylene Chloride      | 4.51  | 84   | 9118     | 3.577   | ug/l    | 90       |
| 43) Isopropyl Acetate       | 8.13  | 43   | 77101    | 11.111  | ug/l    | 99       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101220\  
Data File : VN064087.D  
Acq On : 12 Oct 2020 19:02  
Operator : JC/MD  
Sample : L4225-09  
Misc : 5.00mL/MSVOA N/WATER  
ALS Vial : 12 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
SU-01-100920

Quant Time: Oct 13 02:32:05 2020  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_N\METHODS\82N093020W.M  
Quant Title : SW846 8260  
QLast Update : Thu Oct 01 01:46:38 2020  
Response via : Initial Calibration

