

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101220\
 Data File : VN064089.D
 Acq On : 12 Oct 2020 19:50
 Operator : JC/MD
 Sample : L4333-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OH-WC-C-J4

Manual Integrations
 APPROVED

MMDadoda
 10/13/2020 4:09:53 PM

Quant Time: Oct 13 02:39:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	227304	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	419913	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	431747	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	169305	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	191308	50.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.88%	
35) Dibromofluoromethane	7.55	113	112725	38.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.38%	
50) Toluene-d8	10.06	98	533258	48.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.56%	
62) 4-Bromofluorobenzene	12.38	95	208044	50.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.22%	
Target Compounds						
16) Acetone	3.78	43	37842	32.736	ug/l	96
18) Methyl Acetate	4.29	43	3120m	1.040	ug/l	
20) Methylene Chloride	4.51	84	12098	4.658	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101220\
Data File : VN064089.D
Acq On : 12 Oct 2020 19:50
Operator : JC/MD
Sample : L4333-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OH-WC-C-J4

Manual Integrations
APPROVED

MMDadoda
10/13/2020 4:09:53 PM

Quant Time: Oct 13 02:39:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

