

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101220\
 Data File : VN064091.D
 Acq On : 12 Oct 2020 20:38
 Operator : JC/MD
 Sample : L4333-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OH-WC-C-C1

Quant Time: Oct 13 10:46:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	226286	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	414617	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	434245	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	171406	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	192482	51.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.98%	
35) Dibromofluoromethane	7.55	113	86661	30.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	60.24%	
50) Toluene-d8	10.06	98	537304	49.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.54%	
62) 4-Bromofluorobenzene	12.38	95	207213	50.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.08%	
Target Compounds						
16) Acetone	3.78	43	49342	42.876	ug/l	98
20) Methylene Chloride	4.50	84	12060	4.664	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101220\
Data File : VN064091.D
Acq On : 12 Oct 2020 20:38
Operator : JC/MD
Sample : L4333-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OH-WC-C-C1

Quant Time: Oct 13 10:46:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

