

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN101220\
 Data File : VN064077.D
 Acq On : 12 Oct 2020 14:44
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 13 01:33:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.528	0.425	19.5	83	0.00
3 P	Chloromethane	0.619	0.549	11.3	94	0.00
4 C	Vinyl Chloride	0.621	0.526	15.3#	88	0.00
5 T	Bromomethane	0.391	0.339	13.3	91	0.00
6 T	Chloroethane	0.371	0.347	6.5	94	0.00
7 T	Trichlorofluoromethane	0.986	0.861	12.7	89	0.00
8 T	Diethyl Ether	0.333	0.322	3.3	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.527	0.488	7.4	94	0.00
10 T	Methyl Iodide	0.612	0.569	7.0	88	0.00
11 T	Tert butyl alcohol	0.091	0.109	-19.8	115	0.00
12 CM	1,1-Dichloroethene	0.503	0.449	10.7#	92	0.00
13 T	Acrolein	0.056	0.053	5.4	89	0.00
14 T	Allyl chloride	0.935	0.988	-5.7	105	0.00
15 T	Acrylonitrile	0.261	0.297	-13.8	110	0.00
16 T	Acetone	0.254	0.288	-13.4	113	0.00
17 T	Carbon Disulfide	1.465	1.186	19.0	87	0.00
18 T	Methyl Acetate	0.660	0.801	-21.4	125	0.00
19 T	Methyl tert-butyl Ether	1.829	1.812	0.9	97	0.00
20 T	Methylene Chloride	0.630	0.573	9.0	97	0.00
21 T	trans-1,2-Dichloroethene	0.564	0.504	10.6	92	0.00
22 T	Diisopropyl ether	1.925	2.085	-8.3	106	0.00
23 T	Vinyl Acetate	1.594	1.697	-6.5	101	0.00
24 P	1,1-Dichloroethane	1.100	1.083	1.5	98	0.00
25 T	2-Butanone	0.361	0.431	-19.4	114	0.00
26 T	2,2-Dichloropropane	1.008	0.921	8.6	92	0.00
27 T	cis-1,2-Dichloroethene	0.649	0.615	5.2	95	0.00
28 T	Bromochloromethane	0.544	0.564	-3.7	100	0.00
29 T	Tetrahydrofuran	0.242	0.285	-17.8	114	0.00
30 C	Chloroform	1.147	1.118	2.5#	97	0.00
31 T	Cyclohexane	0.959	0.866	9.7	94	0.00
32 T	1,1,1-Trichloroethane	1.035	0.973	6.0	91	0.00
33 S	1,2-Dichloroethane-d4	0.826	0.728	11.9	87	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
35 S	Dibromofluoromethane	0.347	0.326	6.1	90	0.00
36 T	1,1-Dichloropropene	0.507	0.502	1.0	96	0.00
37 T	Ethyl Acetate	0.503	0.564	-12.1	110	0.00
38 T	Carbon Tetrachloride	0.568	0.522	8.1	90	0.00
39 T	Methylcyclohexane	0.505	0.466	7.7	87	0.00
40 TM	Benzene	1.447	1.476	-2.0	99	0.00
41 T	Methacrylonitrile	0.224	0.241	-7.6	93	0.00
42 TM	1,2-Dichloroethane	0.618	0.579	6.3	90	0.00
43 T	Isopropyl Acetate	0.830	0.924	-11.3	102	0.00
44 TM	Trichloroethene	0.379	0.363	4.2	94	0.00
45 C	1,2-Dichloropropane	0.391	0.417	-6.6#	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN101220\
 Data File : VN064077.D
 Acq On : 12 Oct 2020 14:44
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 13 01:33:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.270	0.265	1.9	95	0.00
47 T	Bromodichloromethane	0.556	0.569	-2.3	96	0.00
48 T	Methyl methacrylate	0.410	0.474	-15.6	108	0.00
49 T	1,4-Dioxane	0.006	0.007	-16.7	116	0.00
50 S	Toluene-d8	1.315	1.235	6.1	91	0.00
51 T	4-Methyl-2-Pentanone	0.487	0.583	-19.7	108	0.00
52 CM	Toluene	0.895	0.917	-2.5#	96	0.00
53 T	t-1,3-Dichloropropene	0.596	0.619	-3.9	96	0.00
54 T	cis-1,3-Dichloropropene	0.623	0.654	-5.0	96	0.00
55 T	1,1,2-Trichloroethane	0.359	0.374	-4.2	97	0.00
56 T	Ethyl methacrylate	0.509	0.568	-11.6	100	0.00
57 T	1,3-Dichloropropane	0.607	0.634	-4.4	96	0.00
58 T	2-Chloroethyl Vinyl ether	0.272	0.309	-13.6	103	0.00
59 T	2-Hexanone	0.362	0.429	-18.5	108	0.00
60 T	Dibromochloromethane	0.418	0.422	-1.0	96	0.00
61 T	1,2-Dibromoethane	0.367	0.378	-3.0	95	0.00
62 S	4-Bromofluorobenzene	0.494	0.453	8.3	88	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	93	0.00
64 T	Tetrachloroethene	0.387	0.343	11.4	90	0.00
65 PM	Chlorobenzene	1.026	1.017	0.9	96	0.00
66 T	1,1,1,2-Tetrachloroethane	0.393	0.400	-1.8	96	0.00
67 C	Ethyl Benzene	1.866	1.891	-1.3#	94	0.00
68 T	m/p-Xylenes	0.696	0.693	0.4	93	0.00
69 T	o-Xylene	0.650	0.667	-2.6	96	0.00
70 T	Styrene	1.104	1.154	-4.5	95	0.00
71 P	Bromoform	0.301	0.309	-2.7	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
73 T	Isopropylbenzene	3.647	3.749	-2.8	92	0.00
74 T	N-amyl acetate	1.509	1.781	-18.0	99	0.00
75 P	1,1,2,2-Tetrachloroethane	1.115	1.211	-8.6	99	0.00
76 T	1,2,3-Trichloropropane	1.102	1.127	-2.3	100	0.00
77 T	Bromobenzene	0.895	0.916	-2.3	94	0.00
78 T	n-propylbenzene	4.244	4.347	-2.4	93	0.00
79 T	2-Chlorotoluene	2.610	2.659	-1.9	95	0.00
80 T	1,3,5-Trimethylbenzene	3.060	3.185	-4.1	93	0.00
81 T	trans-1,4-Dichloro-2-butene	0.383	0.433	-13.1	99	0.00
82 T	4-Chlorotoluene	2.710	2.819	-4.0	94	0.00
83 T	tert-Butylbenzene	2.503	2.535	-1.3	91	0.00
84 T	1,2,4-Trimethylbenzene	3.055	3.175	-3.9	90	0.00
85 T	sec-Butylbenzene	3.287	3.353	-2.0	91	0.00
86 T	p-Isopropyltoluene	2.908	3.003	-3.3	90	0.00
87 T	1,3-Dichlorobenzene	1.619	1.630	-0.7	91	0.00
88 T	1,4-Dichlorobenzene	1.686	1.630	3.3	92	0.00
89 T	n-Butylbenzene	2.684	2.559	4.7	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN101220\
Data File : VN064077.D
Acq On : 12 Oct 2020 14:44
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Oct 13 01:33:29 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.564	0.562	0.4	88	0.00
91 T	1,2-Dichlorobenzene	1.604	1.598	0.4	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.262	0.258	1.5	91	0.00
93 T	1,2,4-Trichlorobenzene	0.881	0.722	18.0	77	0.00
94 T	Hexachlorobutadiene	0.511	0.443	13.3	92	0.00
95 T	Naphthalene	2.699	2.122	21.4	73	0.00
96 T	1,2,3-Trichlorobenzene	0.883	0.738	16.4	81	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6