

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101221\
 Data File : VN068871.D
 Acq On : 12 Oct 2021 19:24
 Operator : JC/MD
 Sample : M4156-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20-20211008-WC

Quant Time: Oct 13 08:22:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	423843	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	687401	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	599335	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	220435	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	252080	45.109	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	90.220%		
35) Dibromofluoromethane	8.024	113	206303	51.584	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	103.160%		
50) Toluene-d8	10.443	98	829325	51.826	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	103.660%		
62) 4-Bromofluorobenzene	12.734	95	287217	47.573	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	95.140%		
Target Compounds					Qvalue	
16) Acetone	4.307	43	5261	2.523	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101221\
Data File : VN068871.D
Acq On : 12 Oct 2021 19:24
Operator : JC/MD
Sample : M4156-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20-20211008-WC

Quant Time: Oct 13 08:22:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

