

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101221\
 Data File : VN068873.D
 Acq On : 12 Oct 2021 20:15
 Operator : JC/MD
 Sample : M4159-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T9-3-BOTTOM-GRAB

Quant Time: Oct 13 08:22:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	428831	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	697904	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	639620	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	237138	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	253869	44.901	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	89.800%
35) Dibromofluoromethane	8.024	113	212014	52.214	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	104.420%
50) Toluene-d8	10.443	98	854225	52.579	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.160%
62) 4-Bromofluorobenzene	12.734	95	286511	46.742	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.480%
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	35901	17.015	ug/l	99
20) Methylene Chloride	5.103	84	21187	4.363	ug/l	98
43) Isopropyl Acetate	8.563	43	44600	4.106	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101221\
Data File : VN068873.D
Acq On : 12 Oct 2021 20:15
Operator : JC/MD
Sample : M4159-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T9-3-BOTTOM-GRAB

Quant Time: Oct 13 08:22:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

