

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN101221\  
 Data File : VN068880.D  
 Acq On : 12 Oct 2021 23:12  
 Operator : JC/MD  
 Sample : M4159-16  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 25 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 T10-2-BOTTOM-GRAB

Quant Time: Oct 13 08:24:50 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N101221W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Oct 13 05:44:40 2021  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) Pentafluorobenzene	8.088	168	425332	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	692039	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	641429	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	243157	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	251582	44.862	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	89.720%
35) Dibromofluoromethane	8.024	113	209655	52.071	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	104.140%
50) Toluene-d8	10.443	98	843158	52.337	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.680%
62) 4-Bromofluorobenzene	12.733	95	283170	46.588	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.180%
Target Compounds						
						Qvalue
16) Acetone	4.293	43	31051	14.837	ug/l	98
20) Methylene Chloride	5.101	84	18277	3.794	ug/l	97
43) Isopropyl Acetate	8.566	43	55161	5.121	ug/l #	86
-----						

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN101221\  
 Data File : VN068880.D  
 Acq On : 12 Oct 2021 23:12  
 Operator : JC/MD  
 Sample : M4159-16  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 25 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 T10-2-BOTTOM-GRAB

Quant Time: Oct 13 08:24:50 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N101221W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Oct 13 05:44:40 2021  
 Response via : Initial Calibration

