

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101221\
 Data File : VN068882.D
 Acq On : 13 Oct 2021 00:02
 Operator : JC/MD
 Sample : M4159-18
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T10-4-BOTTOM-GRAB

Quant Time: Oct 13 08:25:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	487676	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	777385	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	709257	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	270383	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	284420	44.234	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	88.460%		
35) Dibromofluoromethane	8.024	113	234849	51.924	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	103.840%		
50) Toluene-d8	10.443	98	935567	51.698	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	103.400%		
62) 4-Bromofluorobenzene	12.733	95	321394	47.072	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	94.140%		
Target Compounds					Qvalue	
16) Acetone	4.296	43	31829	13.265	ug/l	99
20) Methylene Chloride	5.100	84	18268	3.308	ug/l	96
43) Isopropyl Acetate	8.568	43	64297	5.314	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101221\
Data File : VN068882.D
Acq On : 13 Oct 2021 00:02
Operator : JC/MD
Sample : M4159-18
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T10-4-BOTTOM-GRAB

Quant Time: Oct 13 08:25:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

