

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN101221\  
 Data File : VN068883.D  
 Acq On : 13 Oct 2021 00:27  
 Operator : JC/MD  
 Sample : M4159-04  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 28 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 T9-4-TOP-GRAB

Quant Time: Oct 13 08:25:45 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N101221W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Oct 13 05:44:40 2021  
 Response via : Initial Calibration


| Compound                    | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|--------|----------|
| -----                       |        |       |          |          |        |          |
| Internal Standards          |        |       |          |          |        |          |
| 1) Pentafluorobenzene       | 8.088  | 168   | 427180   | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 8.971  | 114   | 698078   | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 11.746 | 117   | 643553   | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.677 | 152   | 251629   | 50.000   | ug/l   | 0.00     |
| System Monitoring Compounds |        |       |          |          |        |          |
| 33) 1,2-Dichloroethane-d4   | 8.440  | 65    | 255401   | 45.346   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 61 - 141 | Recovery | =      | 90.700%  |
| 35) Dibromofluoromethane    | 8.024  | 113   | 209802   | 51.657   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 69 - 133 | Recovery | =      | 103.320% |
| 50) Toluene-d8              | 10.443 | 98    | 844073   | 51.941   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 65 - 126 | Recovery | =      | 103.880% |
| 62) 4-Bromofluorobenzene    | 12.733 | 95    | 295858   | 48.255   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 58 - 135 | Recovery | =      | 96.500%  |
| Target Compounds            |        |       |          |          |        |          |
|                             |        |       |          |          | Qvalue |          |
| 16) Acetone                 | 4.291  | 43    | 31615    | 15.042   | ug/l   | 99       |
| 18) Methyl Acetate          | 4.867  | 43    | 11324    | 1.420    | ug/l   | 98       |
| 20) Methylene Chloride      | 5.106  | 84    | 17518    | 3.621    | ug/l   | 92       |
| 43) Isopropyl Acetate       | 8.566  | 43    | 84127    | 7.743    | ug/l # | 77       |
| -----                       |        |       |          |          |        |          |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN101221\  
Data File : VN068883.D  
Acq On : 13 Oct 2021 00:27  
Operator : JC/MD  
Sample : M4159-04  
Misc : 5.00mL/MSVOA\_N/WATER  
ALS Vial : 28 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
T9-4-TOP-GRAB

Quant Time: Oct 13 08:25:45 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N101221W.M  
Quant Title : SW846 8260  
QLast Update : Wed Oct 13 05:44:40 2021  
Response via : Initial Calibration









#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 8.088 min Scan# 2274  
Delta R.T. 0.002 min  
Lab File: VN068883.D  
Acq: 13 Oct 2021 00:27




Instrument :  
MSVOA\_N  
ClientSampleId :  
T9-4-TOP-GRAB


Tgt Ion:168 Resp: 427180  
Ion Ratio Lower Upper  
168 100  
99 58.4 48.9 73.3



#16  
Acetone  
Concen: 15.042 ug/l  
RT: 4.291 min Scan# 858  
Delta R.T. 0.005 min  
Lab File: VN068883.D  
Acq: 13 Oct 2021 00:27

Tgt Ion: 43 Resp: 31615  
Ion Ratio Lower Upper  
43 100  
58 29.9 24.5 36.7





#18

Methyl Acetate

Concen: 1.420 ug/l

RT: 4.867 min Scan# 1073

Delta R.T. 0.013 min

Lab File: VN068883.D

Acq: 13 Oct 2021 00:27

Tgt Ion: 43 Resp: 11324

Ion Ratio Lower Upper

43 100





74 24.5 18.9 28.3

Instrument :

MSVOA\_N

ClientSampleId :

T9-4-TOP-GRAB



#20

Methylene Chloride

Concen: 3.621 ug/l

RT: 5.106 min Scan# 1162

Delta R.T. 0.011 min

Lab File: VN068883.D

Acq: 13 Oct 2021 00:27

Tgt Ion: 84 Resp: 17518




Ion Ratio Lower Upper


84 100

49 110.7 97.6 146.4

51 31.3 30.3 45.5

86 62.1 51.8 77.6









#33

1,2-Dichloroethane-d4  
Concen: 45.346 ug/l  
RT: 8.440 min Scan# 2405  
Delta R.T. -0.000 min  
Lab File: VN068883.D  
Acq: 13 Oct 2021 00:27

Instrument :  
MSVOA\_N  
ClientSampleId :  
T9-4-TOP-GRAB




Tgt Ion: 65 Resp: 255401  
Ion Ratio Lower Upper  
65 100  
67 52.2 0.0 103.8




#34

1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 8.971 min Scan# 2603  
Delta R.T. 0.003 min  
Lab File: VN068883.D  
Acq: 13 Oct 2021 00:27

Tgt Ion: 114 Resp: 698078  
Ion Ratio Lower Upper  
114 100  
63 20.8 0.0 42.6  
88 15.9 0.0 32.0









#35  
Dibromofluoromethane  
Concen: 51.657 ug/l  
RT: 8.024 min Scan# 2250  
Delta R.T. 0.003 min  
Lab File: VN068883.D  
Acq: 13 Oct 2021 00:27




Instrument :  
MSVOA\_N  
ClientSampleId :  
T9-4-TOP-GRAB


Tgt Ion:113 Resp: 209802  
Ion Ratio Lower Upper  
113 100  
111 102.1 83.2 124.8  
192 19.5 15.0 22.4



#43  
Isopropyl Acetate  
Concen: 7.743 ug/l  
RT: 8.566 min Scan# 2452  
Delta R.T. 0.005 min  
Lab File: VN068883.D  
Acq: 13 Oct 2021 00:27

Tgt Ion: 43 Resp: 84127  
Ion Ratio Lower Upper  
43 100  
61 11.7 20.4 30.6#  
87 7.0 10.7 16.1#









#50  
Toluene-d8  
Concen: 51.941 ug/l  
RT: 10.443 min Scan# 3152  
Delta R.T. -0.000 min  
Lab File: VN068883.D  
Acq: 13 Oct 2021 00:27




Instrument :  
MSVOA\_N  
ClientSampleId :  
T9-4-TOP-GRAB


Tgt Ion: 98 Resp: 844073  
Ion Ratio Lower Upper  
98 100  
100 64.3 51.5 77.3



#62  
4-Bromofluorobenzene  
Concen: 48.255 ug/l  
RT: 12.733 min Scan# 4006  
Delta R.T. -0.000 min  
Lab File: VN068883.D  
Acq: 13 Oct 2021 00:27

Tgt Ion: 95 Resp: 295858  
Ion Ratio Lower Upper  
95 100  
174 77.3 0.0 144.8  
176 74.8 0.0 140.0









#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.746 min Scan# 3638  
Delta R.T. -0.000 min  
Lab File: VN068883.D  
Acq: 13 Oct 2021 00:27

Instrument :  
MSVOA\_N  
ClientSampleId :  
T9-4-TOP-GRAB

Tgt Ion:117 Resp: 643553  
Ion Ratio Lower Upper  
117 100  
82 56.4 46.4 69.6  
119 31.8 26.0 39.0



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 13.677 min Scan# 4358  
Delta R.T. 0.003 min  
Lab File: VN068883.D  
Acq: 13 Oct 2021 00:27

Tgt Ion:152 Resp: 251629  
Ion Ratio Lower Upper  
152 100  
115 60.7 31.4 94.3  
150 157.0 0.0 351.0

