

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101222\
 Data File : VN074875.D
 Acq On : 12 Oct 2022 14:29
 Operator : JC\MD
 Sample : VN1012WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1012WBL01

Quant Time: Oct 13 07:55:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N101222W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Oct 13 07:52:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.824	128	66456	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.106	114	376850	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.871	117	313500	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.588	65	197609	31.761	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	105.867%
60) 4-Bromofluorobenzene	12.847	95	139477	25.088	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	83.633%
63) Toluene-d8	10.571	98	582356	32.186	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	107.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101222\
Data File : VN074875.D
Acq On : 12 Oct 2022 14:29
Operator : JC\MD
Sample : VN1012WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1012WBL01

Quant Time: Oct 13 07:55:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N101222W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Oct 13 07:52:24 2022
Response via : Initial Calibration

