

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101222\
 Data File : VN074877.D
 Acq On : 12 Oct 2022 15:42
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Manual Integrations
 APPROVED

Reviewed By : John Carlone 10/13/2022
 Supervised By : Mahesh Dadoda 10/14/2022

Quant Time: Oct 13 07:56:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N101222W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Oct 13 07:52:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.824	128	67584	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	0.000	114	0m	30.000	ug/l	-9.11
57) Chlorobenzene-d5	11.871	117	323548	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.583	65	198089m	31.307	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	104.367%
60) 4-Bromofluorobenzene	12.853	95	139777	24.361	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	81.200%
63) Toluene-d8	10.571	98	573946	30.736	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	102.467%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101222\
Data File : VN074877.D
Acq On : 12 Oct 2022 15:42
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/13/2022
Supervised By :Mahesh Dadoda 10/14/2022

Quant Time: Oct 13 07:56:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N101222W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Oct 13 07:52:24 2022
Response via : Initial Calibration

