

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101222\
 Data File : VN074886.D
 Acq On : 12 Oct 2022 19:34
 Operator : JC\MD
 Sample : N5089-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 001-WILLETS-PT-BLVD(OCT)

Quant Time: Oct 13 08:03:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N101222W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Oct 13 07:52:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.824	128	67430	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.112	114	378566	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.871	117	329460	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.594	65	193217	30.607	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	102.033%
60) 4-Bromofluorobenzene	12.853	95	143447	24.552	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	81.833%
63) Toluene-d8	10.571	98	585939	30.815	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	102.700%
Target Compounds						
15) Acetone	4.454	58	331736	327.192	ug/l	Qvalue 87
30) 2-Butanone	7.506	43	40615	7.452	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101222\
Data File : VN074886.D
Acq On : 12 Oct 2022 19:34
Operator : JC\MD
Sample : N5089-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
001-WILLETS-PT-BLVD(OCT)

Quant Time: Oct 13 08:03:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N101222W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Oct 13 07:52:24 2022
Response via : Initial Calibration

