

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101222\
 Data File : VN074887.D
 Acq On : 12 Oct 2022 19:58
 Operator : JC\MD
 Sample : N5089-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 002-35TH-AVE(OCT)

Manual Integrations
 APPROVED

Reviewed By : John Carlone 10/13/2022
 Supervised By : Mahesh Dadoda 10/14/2022

Quant Time: Oct 13 08:03:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N101222W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Oct 13 07:52:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.830	128	60236	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.106	114	361419m	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.865	117	304399	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.588	65	193585	34.327	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	114.433%#
60) 4-Bromofluorobenzene	12.847	95	142284	26.358	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	87.867%
63) Toluene-d8	10.571	98	543452	30.934	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	103.100%
Target Compounds						
15) Acetone	4.453	58	335065	369.944	ug/l	Qvalue 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101222\
Data File : VN074887.D
Acq On : 12 Oct 2022 19:58
Operator : JC\MD
Sample : N5089-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
002-35TH-AVE(OCT)

Quant Time: Oct 13 08:03:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N101222W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Oct 13 07:52:24 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/13/2022
Supervised By :Mahesh Dadoda 10/14/2022

