

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101223\
 Data File : VN079544.D
 Acq On : 12 Oct 2023 14:47
 Operator : JC\MD
 Sample : 04844-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-603-COMP-06

Quant Time: Oct 13 01:19:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	261632	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	525466	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	524778	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	169605	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	218324	51.465	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 102.940%		
35) Dibromofluoromethane	8.171	113	193304	51.190	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 102.380%		
50) Toluene-d8	10.571	98	651167	47.658	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 95.320%		
62) 4-Bromofluorobenzene	12.853	95	215891	49.236	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 98.480%		
Target Compounds						
						Qvalue
16) Acetone	4.430	43	34775	20.668	ug/l	94
20) Methylene Chloride	5.283	84	15410	3.624	ug/l #	90
43) Isopropyl Acetate	8.694	43	236984	23.986	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101223\
Data File : VN079544.D
Acq On : 12 Oct 2023 14:47
Operator : JC\MD
Sample : 04844-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-603-COMP-06

Quant Time: Oct 13 01:19:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

