

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101223\
 Data File : VN079545.D
 Acq On : 12 Oct 2023 15:11
 Operator : JC\MD
 Sample : 04844-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-603-COMP-07

Quant Time: Oct 13 01:20:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	248090	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	509969	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	511471	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	164752	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	212205	52.753	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.500%
35) Dibromofluoromethane	8.171	113	184995	50.479	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.960%
50) Toluene-d8	10.571	98	629385	47.464	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.920%
62) 4-Bromofluorobenzene	12.853	95	211773	49.764	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.520%
Target Compounds						
						Qvalue
16) Acetone	4.442	43	35569	22.293	ug/l #	84
20) Methylene Chloride	5.277	84	15105	3.767	ug/l #	80
43) Isopropyl Acetate	8.694	43	233760	24.379	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101223\
Data File : VN079545.D
Acq On : 12 Oct 2023 15:11
Operator : JC\MD
Sample : 04844-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-603-COMP-07

Quant Time: Oct 13 01:20:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

