

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101223\
 Data File : VN079546.D
 Acq On : 12 Oct 2023 15:35
 Operator : JC\MD
 Sample : 04844-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-603-COMP-08

Quant Time: Oct 13 01:20:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	245684	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	499200	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	513462	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	160757	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	210300	52.792	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 105.580%	
35) Dibromofluoromethane	8.171	113	182983	51.007	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.020%	
50) Toluene-d8	10.571	98	626898	48.296	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 96.600%	
62) 4-Bromofluorobenzene	12.853	95	207433	49.796	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.600%	
Target Compounds						
						Qvalue
16) Acetone	4.436	43	35754	22.629	ug/l	95
20) Methylene Chloride	5.289	84	15061	3.797	ug/l #	84
43) Isopropyl Acetate	8.694	43	234153	24.947	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101223\
Data File : VN079546.D
Acq On : 12 Oct 2023 15:35
Operator : JC\MD
Sample : 04844-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-603-COMP-08

Quant Time: Oct 13 01:20:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

