

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101223\
 Data File : VN079547.D
 Acq On : 12 Oct 2023 15:59
 Operator : JC\MD
 Sample : 04858-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-1

Quant Time: Oct 13 01:21:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	252860	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	511717	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	508827	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	155628	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	215855	52.648	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.300%
35) Dibromofluoromethane	8.171	113	189452	51.518	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.040%
50) Toluene-d8	10.571	98	624824	46.959	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.920%
62) 4-Bromofluorobenzene	12.853	95	204182	47.816	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.640%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	61883	38.054	ug/l	94
20) Methylene Chloride	5.277	84	17606	4.395	ug/l #	89
25) 2-Butanone	7.488	43	17479	7.242	ug/l #	73
43) Isopropyl Acetate	8.694	43	198260	20.606	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101223\
Data File : VN079547.D
Acq On : 12 Oct 2023 15:59
Operator : JC\MD
Sample : 04858-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-1

Quant Time: Oct 13 01:21:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

