

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101223\
 Data File : VN079548.D
 Acq On : 12 Oct 2023 16:23
 Operator : JC\MD
 Sample : 04858-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-2

Quant Time: Oct 13 01:21:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	246946	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	515352	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	508131	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	157414	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	215630	53.853	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.700%	
35) Dibromofluoromethane	8.171	113	187148	50.533	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.060%	
50) Toluene-d8	10.571	98	624607	46.611	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 93.220%	
62) 4-Bromofluorobenzene	12.853	95	209684	48.759	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.520%	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	41217	25.953	ug/l #	87
20) Methylene Chloride	5.289	84	16959	4.326	ug/l #	78
43) Isopropyl Acetate	8.694	43	214592	22.146	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101223\
Data File : VN079548.D
Acq On : 12 Oct 2023 16:23
Operator : JC\MD
Sample : 04858-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-2

Quant Time: Oct 13 01:21:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

