

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101223\
 Data File : VN079551.D
 Acq On : 12 Oct 2023 17:35
 Operator : JC\MD
 Sample : PB156319ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB156319ZHE#02

Quant Time: Oct 13 01:23:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	175507	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	377502	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	374904	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	101457	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	158799	55.803	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 111.600%			
35) Dibromofluoromethane	8.171	113	132062	48.680	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 97.360%			
50) Toluene-d8	10.571	98	500537	50.992	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.980%			
62) 4-Bromofluorobenzene	12.853	95	137553	43.666	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 87.340%			
Target Compounds						
16) Acetone	4.436	43	30745	27.239	ug/l	91
20) Methylene Chloride	5.283	84	20171	7.651	ug/l	92
43) Isopropyl Acetate	8.694	43	283542	39.947	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101223\
Data File : VN079551.D
Acq On : 12 Oct 2023 17:35
Operator : JC\MD
Sample : PB156319ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB156319ZHE#02

Quant Time: Oct 13 01:23:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

