

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101223\
 Data File : VN079554.D
 Acq On : 12 Oct 2023 18:47
 Operator : JC\MD
 Sample : PB156319ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB156319ZHE#05

Quant Time: Oct 13 01:24:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	176900	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	380547	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	383120	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	105209	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	160203	55.853	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	=	111.700%	
35) Dibromofluoromethane	8.171	113	131104	47.940	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	=	95.880%	
50) Toluene-d8	10.571	98	500849	50.616	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	=	101.240%	
62) 4-Bromofluorobenzene	12.853	95	143995	45.345	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery	=	90.680%	
Target Compounds						Qvalue
16) Acetone	4.436	43	30658	26.948	ug/l	93
20) Methylene Chloride	5.289	84	19613	7.359	ug/l #	77
43) Isopropyl Acetate	8.694	43	301446	42.130	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101223\
Data File : VN079554.D
Acq On : 12 Oct 2023 18:47
Operator : JC\MD
Sample : PB156319ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB156319ZHE#05

Quant Time: Oct 13 01:24:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

