

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101223\
 Data File : VN079555.D
 Acq On : 12 Oct 2023 19:11
 Operator : JC\MD
 Sample : PB156319ZHE#06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB156319ZHE#06

Quant Time: Oct 13 01:25:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	183459	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	398691	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	387704	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	99127	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	164622	55.342	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.680%
35) Dibromofluoromethane	8.171	113	136955	47.800	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.600%
50) Toluene-d8	10.570	98	525539	50.694	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.380%
62) 4-Bromofluorobenzene	12.853	95	142229	42.751	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.500%
Target Compounds						
						Qvalue
16) Acetone	4.436	43	30510	25.859	ug/l	89
20) Methylene Chloride	5.283	84	19653	7.090	ug/l #	87
43) Isopropyl Acetate	8.694	43	310278	41.391	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101223\
Data File : VN079555.D
Acq On : 12 Oct 2023 19:11
Operator : JC\MD
Sample : PB156319ZHE#06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB156319ZHE#06

Quant Time: Oct 13 01:25:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

