

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101319\
 Data File : VN058711.D
 Acq On : 11 Oct 2019 18:55
 Operator : JC/SP
 Sample : K5349-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Oct 12 00:52:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 12 00:00:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	426308	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	674453	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	593218	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	243306	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	306121	51.29	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.58%	
35) Dibromofluoromethane	7.57	113	219083	52.59	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.18%	
50) Toluene-d8	10.08	98	854092	51.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.54%	
62) 4-Bromofluorobenzene	12.41	95	283871	45.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.04%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101319\
Data File : VN058711.D
Acq On : 11 Oct 2019 18:55
Operator : JC/SP
Sample : K5349-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Oct 12 00:52:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 12 00:00:30 2019
Response via : Initial Calibration

