

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101319\
 Data File : VN058713.D
 Acq On : 11 Oct 2019 19:39
 Operator : JC/SP
 Sample : K5349-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: Oct 12 00:55:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 12 00:00:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	412713	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	658178	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	572272	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	232564	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	299792	51.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.78%	
35) Dibromofluoromethane	7.57	113	213308	52.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.94%	
50) Toluene-d8	10.08	98	823369	50.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.30%	
62) 4-Bromofluorobenzene	12.41	95	272353	44.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.52%	
Target Compounds						
16) Acetone	3.80	43	13125	5.487	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101319\
Data File : VN058713.D
Acq On : 11 Oct 2019 19:39
Operator : JC/SP
Sample : K5349-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: Oct 12 00:55:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 12 00:00:30 2019
Response via : Initial Calibration

