

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101319\
 Data File : VN058718.D
 Acq On : 11 Oct 2019 21:30
 Operator : JC/SP
 Sample : K5336-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FA19-JMW6001

Quant Time: Oct 12 01:07:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 12 00:00:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	422099	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	661634	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	585943	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	232529	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.01	65	289359	48.97	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	97.94%	
35) Dibromofluoromethane	7.57	113	207329	50.73	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	101.46%	
50) Toluene-d8	10.08	98	806131	49.33	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	98.66%	
62) 4-Bromofluorobenzene	12.41	95	262341	42.41	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	84.82%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	2.04	50	2586	0.414	ug/l	94
4) Vinyl Chloride	2.17	62	9233	1.483	ug/l	94
16) Acetone	3.81	43	15179	6.205	ug/l	97
21) trans-1,2-Dichloroethene	5.02	96	2629	0.580	ug/l #	81
27) cis-1,2-Dichloroethene	6.81	96	411515	77.827	ug/l	87
37) Ethyl Acetate	6.91	43	16808	2.484	ug/l #	74
44) Trichloroethene	8.82	130	202261	42.393	ug/l	97
64) Tetrachloroethene	10.63	164	1265	0.308	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101319\
Data File : VN058718.D
Acq On : 11 Oct 2019 21:30
Operator : JC/SP
Sample : K5336-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FA19-JMW6001

Quant Time: Oct 12 01:07:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 12 00:00:30 2019
Response via : Initial Calibration

