

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101319\
 Data File : VN058724.D
 Acq On : 12 Oct 2019 00:27
 Operator : JC/SP
 Sample : PB123806TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB123806TB

Manual Integrations
 APPROVED

MMDadoda
 10/16/2019 12:48:57 AM

Quant Time: Oct 12 01:19:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 12 00:00:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	416379	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	655029	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	593727	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	236512	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	294959	50.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.20%	
35) Dibromofluoromethane	7.57	113	210873	52.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.24%	
50) Toluene-d8	10.08	98	845101	52.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.46%	
62) 4-Bromofluorobenzene	12.41	95	275812	45.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.08%	
Target Compounds						
16) Acetone	3.80	43	12124m	5.024	ug/l	
20) Methylene Chloride	4.53	84	22615	4.484	ug/l	93
43) Isopropyl Acetate	8.15	43	108456	9.505	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101319\
Data File : VN058724.D
Acq On : 12 Oct 2019 00:27
Operator : JC/SP
Sample : PB123806TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB123806TB

Manual Integrations
APPROVED

MMDadoda
10/16/2019 12:48:57 AM

Quant Time: Oct 12 01:19:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 12 00:00:30 2019
Response via : Initial Calibration

