

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101319\
 Data File : VN058741.D
 Acq On : 12 Oct 2019 6:40
 Operator : JC/SP
 Sample : K5266-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AMENDED-COMPOSITE-C

Quant Time: Oct 12 07:10:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 12 00:00:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	421618	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	660629	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	585323	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	235393	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	303747	51.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.92%	
35) Dibromofluoromethane	7.57	113	88021	21.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	43.14%	
50) Toluene-d8	10.09	98	836939	51.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.58%	
62) 4-Bromofluorobenzene	12.41	95	275008	44.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.04%	
Target Compounds						
4) Vinyl Chloride	2.17	62	45496	7.316	ug/l	98
16) Acetone	3.80	43	36713	15.024	ug/l	100
20) Methylene Chloride	4.53	84	65863	12.896	ug/l	97
24) 1,1-Dichloroethane	5.83	63	22495	2.486	ug/l	98
27) cis-1,2-Dichloroethene	6.81	96	161683	30.613	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101319\
Data File : VN058741.D
Acq On : 12 Oct 2019 6:40
Operator : JC/SP
Sample : K5266-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AMENDED-COMPOSITE-C

Quant Time: Oct 12 07:10:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 12 00:00:30 2019
Response via : Initial Calibration

