

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101320\
 Data File : VN064105.D
 Acq On : 13 Oct 2020 16:46
 Operator : JC/MD
 Sample : L4217-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HR-01-101220

Manual Integrations
 APPROVED

MMDadoda
 10/14/2020 2:09:16 PM

Quant Time: Oct 14 04:35:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	226856	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	420104	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	450130	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	178125	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	190693	50.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.76%	
35) Dibromofluoromethane	7.55	113	143585	49.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.52%	
50) Toluene-d8	10.06	98	553503	50.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.18%	
62) 4-Bromofluorobenzene	12.38	95	219217	52.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.54%	
Target Compounds						
16) Acetone	3.78	43	16086	13.943	ug/l #	89
18) Methyl Acetate	4.30	43	3740m	1.249	ug/l	
20) Methylene Chloride	4.50	84	11499	4.441	ug/l	86
43) Isopropyl Acetate	8.13	43	81211	11.644	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101320\
Data File : VN064105.D
Acq On : 13 Oct 2020 16:46
Operator : JC/MD
Sample : L4217-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HR-01-101220

Manual Integrations
APPROVED

MMDadoda
10/14/2020 2:09:16 PM

Quant Time: Oct 14 04:35:11 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

