

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101320\
 Data File : VN064106.D
 Acq On : 13 Oct 2020 17:10
 Operator : JC/MD
 Sample : L4333-12RE
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OH-WC-C-ClRE

Quant Time: Oct 14 04:38:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	228979	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	425586	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	442206	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	166867	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	191680	50.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.34%	
35) Dibromofluoromethane	7.55	113	94805	32.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	64.20%	
50) Toluene-d8	10.06	98	542437	48.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.90%	
62) 4-Bromofluorobenzene	12.38	95	213522	50.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.48%	
Target Compounds						
16) Acetone	3.78	43	44267	38.014	ug/l	99
20) Methylene Chloride	4.50	84	12091	4.622	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101320\
Data File : VN064106.D
Acq On : 13 Oct 2020 17:10
Operator : JC/MD
Sample : L4333-12RE
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OH-WC-C-C1RE

Quant Time: Oct 14 04:38:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

