

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101320\
 Data File : VN064107.D
 Acq On : 13 Oct 2020 17:34
 Operator : JC/MD
 Sample : L4333-15RE
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OH-WC-C-D4RE

Quant Time: Oct 14 04:46:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	224429	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	412392	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	426969	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	165355	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	183532	49.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.00%	
35) Dibromofluoromethane	7.55	113	97624	34.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	68.24%	
50) Toluene-d8	10.06	98	530073	48.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.74%	
62) 4-Bromofluorobenzene	12.38	95	203637	49.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.88%	
Target Compounds						
16) Acetone	3.78	43	39308	34.440	ug/l	98
20) Methylene Chloride	4.51	84	12534	4.883	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101320\
Data File : VN064107.D
Acq On : 13 Oct 2020 17:34
Operator : JC/MD
Sample : L4333-15RE
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OH-WC-C-D4RE

Quant Time: Oct 14 04:46:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

