

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101320\
 Data File : VN064109.D
 Acq On : 13 Oct 2020 18:22
 Operator : JC/MD
 Sample : L4381-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-FRAC-TANKS

Manual Integrations
 APPROVED

MMDadoda
 10/14/2020 2:09:20 PM

Quant Time: Oct 14 04:51:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	218568	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	401825	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	403829	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	161091	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	180911	50.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.20%	
35) Dibromofluoromethane	7.55	113	133495	47.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.76%	
50) Toluene-d8	10.06	98	509911	48.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.48%	
62) 4-Bromofluorobenzene	12.38	95	193895	48.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.60%	
Target Compounds						
16) Acetone	3.79	43	9823m	8.837	ug/l	Qvalue
25) 2-Butanone	6.80	43	6100	3.871	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101320\
Data File : VN064109.D
Acq On : 13 Oct 2020 18:22
Operator : JC/MD
Sample : L4381-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-FRAC-TANKS

Manual Integrations
APPROVED

MMDadoda
10/14/2020 2:09:20 PM

Quant Time: Oct 14 04:51:20 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

