

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101321\
 Data File : VN068889.D
 Acq On : 13 Oct 2021 11:19
 Operator : JC/MD
 Sample : VN1013MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1013MBL01

Quant Time: Oct 14 06:40:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	440054	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	715963	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	625533	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	235939	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	258261	44.51	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	89.02%
35) Dibromofluoromethane	8.024	113	214682	51.54	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.08%
50) Toluene-d8	10.443	98	857230	51.43	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.86%
62) 4-Bromofluorobenzene	12.734	95	282698	44.96	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.92%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101321\
Data File : VN068889.D
Acq On : 13 Oct 2021 11:19
Operator : JC/MD
Sample : VN1013MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1013MBL01

Quant Time: Oct 14 06:40:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

