

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101321\
 Data File : VN068913.D
 Acq On : 13 Oct 2021 21:50
 Operator : JC/MD
 Sample : M4192-21 10X
 Misc : 4.68g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HAR-21

Quant Time: Oct 14 06:45:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	414954	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	668031	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	615838	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	238671	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	238638	43.62	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	87.24%
35) Dibromofluoromethane	8.024	113	196794	50.63	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.26%
50) Toluene-d8	10.443	98	815651	52.45	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.90%
62) 4-Bromofluorobenzene	12.733	95	279653	47.66	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.32%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101321\
Data File : VN068913.D
Acq On : 13 Oct 2021 21:50
Operator : JC/MD
Sample : M4192-21 10X
Misc : 4.68g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HAR-21

Quant Time: Oct 14 06:45:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 8.088 min Scan# 2274
Delta R.T. 0.002 min
Lab File: VN068913.D
Acq: 13 Oct 2021 21:50




Instrument :
MSVOA_N
ClientSampled :
HAR-21


Tgt Ion:168 Resp: 414954
Ion Ratio Lower Upper
168 100
99 57.6 48.9 73.3



#33
1,2-Dichloroethane-d4
Concen: 43.62 ug/l
RT: 8.437 min Scan# 2404
Delta R.T. -0.003 min
Lab File: VN068913.D
Acq: 13 Oct 2021 21:50

Tgt Ion: 65 Resp: 238638
Ion Ratio Lower Upper
65 100
67 52.8 0.0 103.8









#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.968 min Scan# 2602
Delta R.T. -0.000 min
Lab File: VN068913.D
Acq: 13 Oct 2021 21:50




Instrument :
MSVOA_N
ClientSampleId :
HAR-21


Tgt Ion:114 Resp: 668031
Ion Ratio Lower Upper
114 100
63 20.2 0.0 42.6
88 15.6 0.0 32.0



#35
Dibromofluoromethane
Concen: 50.63 ug/l
RT: 8.024 min Scan# 2250
Delta R.T. 0.003 min
Lab File: VN068913.D
Acq: 13 Oct 2021 21:50

Tgt Ion:113 Resp: 196794
Ion Ratio Lower Upper
113 100
111 102.8 83.2 124.8
192 19.7 15.0 22.4









#50
Toluene-d8
Concen: 52.45 ug/l
RT: 10.443 min Scan# 3152
Delta R.T. 0.000 min
Lab File: VN068913.D
Acq: 13 Oct 2021 21:50




Instrument :
MSVOA_N
ClientSampled :
HAR-21


Tgt Ion: 98 Resp: 815651
Ion Ratio Lower Upper
98 100
100 63.9 51.5 77.3



#62
4-Bromofluorobenzene
Concen: 47.66 ug/l
RT: 12.733 min Scan# 4006
Delta R.T. 0.000 min
Lab File: VN068913.D
Acq: 13 Oct 2021 21:50

Tgt Ion: 95 Resp: 279653
Ion Ratio Lower Upper
95 100
174 78.7 0.0 144.8
176 76.7 0.0 140.0









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.746 min Scan# 3638
Delta R.T. 0.000 min
Lab File: VN068913.D
Acq: 13 Oct 2021 21:50

Instrument :
MSVOA_N
ClientSampled :
HAR-21

Tgt Ion:117 Resp: 615838
Ion Ratio Lower Upper
117 100
82 55.6 46.4 69.6
119 32.3 26.0 39.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.678 min Scan# 4358
Delta R.T. 0.003 min
Lab File: VN068913.D
Acq: 13 Oct 2021 21:50

Tgt Ion:152 Resp: 238671
Ion Ratio Lower Upper
152 100
115 59.8 31.4 94.3
150 157.7 0.0 351.0

