

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101322\
 Data File : VN074892.D
 Acq On : 13 Oct 2022 12:50
 Operator : JC\MD
 Sample : VN1013WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1013WBL01

Quant Time: Oct 14 05:04:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N101222W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Oct 13 07:52:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.830	128	69874	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.112	114	396097	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.871	117	334445	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.588	65	207844	31.772	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	105.900%
60) 4-Bromofluorobenzene	12.853	95	147618	24.889	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	82.967%
63) Toluene-d8	10.571	98	584052	30.258	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.867%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101322\
Data File : VN074892.D
Acq On : 13 Oct 2022 12:50
Operator : JC\MD
Sample : VN1013WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1013WBL01

Quant Time: Oct 14 05:04:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N101222W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Oct 13 07:52:24 2022
Response via : Initial Calibration

