

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101323\
 Data File : VN079563.D
 Acq On : 13 Oct 2023 11:00
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 13 23:28:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	92	0.00
2 T	Dichlorodifluoromethane	0.716	0.721	-0.7	89	0.00
3 P	Chloromethane	0.894	0.808	9.6	85	0.00
4 C	Vinyl Chloride	0.930	0.925	0.5#	91	0.00
5 T	Bromomethane	0.557	0.602	-8.1	107	0.00
6 T	Chloroethane	0.701	0.613	12.6	88	0.00
7 T	Trichlorofluoromethane	1.098	1.104	-0.5	95	0.00
8 T	Diethyl Ether	0.391	0.402	-2.8	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.642	0.645	-0.5	99	0.00
10 T	Methyl Iodide	0.557	0.718	-28.9#	103	0.00
11 T	Tert butyl alcohol	0.122	0.106	13.1	81	-0.01
12 CM	1,1-Dichloroethene	0.591	0.597	-1.0#	94	0.00
13 T	Acrolein	0.121	0.146	-20.7	108	0.00
14 T	Allyl chloride	0.953	0.936	1.8	92	0.00
15 T	Acrylonitrile	0.354	0.353	0.3	89	0.00
16 T	Acetone	0.322	0.327	-1.6	92	0.00
17 T	Carbon Disulfide	1.843	1.651	10.4	86	0.00
18 T	Methyl Acetate	1.574	1.441	8.4	88	0.00
19 T	Methyl tert-butyl Ether	1.919	2.052	-6.9	94	0.00
20 T	Methylene Chloride	0.816	0.720	11.8	95	0.00
21 T	trans-1,2-Dichloroethene	0.690	0.655	5.1	92	0.00
22 T	Diisopropyl ether	2.138	2.313	-8.2	96	0.00
23 T	Vinyl Acetate	1.202	1.297	-7.9	90	0.00
24 P	1,1-Dichloroethane	1.308	1.318	-0.8	95	0.00
25 T	2-Butanone	0.477	0.479	-0.4	87	0.00
26 T	2,2-Dichloropropane	1.000	1.095	-9.5	98	0.00
27 T	cis-1,2-Dichloroethene	0.750	0.774	-3.2	96	0.00
28 T	Bromochloromethane	0.622	0.533	14.3	75	0.00
29 T	Tetrahydrofuran	0.309	0.309	0.0	86	0.00
30 C	Chloroform	1.322	1.347	-1.9#	96	0.00
31 T	Cyclohexane	1.150	1.074	6.6	92	0.00
32 T	1,1,1-Trichloroethane	1.112	1.143	-2.8	95	0.00
33 S	1,2-Dichloroethane-d4	0.811	0.824	-1.6	96	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	93	0.00
35 S	Dibromofluoromethane	0.359	0.372	-3.6	100	0.00
36 T	1,1-Dichloropropene	0.539	0.563	-4.5	94	0.00
37 T	Ethyl Acetate	0.537	0.548	-2.0	89	0.00
38 T	Carbon Tetrachloride	0.540	0.558	-3.3	95	0.00
39 T	Methylcyclohexane	0.518	0.527	-1.7	87	0.00
40 TM	Benzene	1.602	1.659	-3.6	95	0.00
41 T	Methacrylonitrile	0.282	0.281	0.4	85	0.00
42 TM	1,2-Dichloroethane	0.582	0.593	-1.9	96	0.00
43 T	Isopropyl Acetate	0.940	0.945	-0.5	93	0.00
44 TM	Trichloroethene	0.377	0.381	-1.1	95	0.00
45 C	1,2-Dichloropropane	0.422	0.443	-5.0#	98	0.00
46 T	Dibromomethane	0.273	0.285	-4.4	94	0.00
47 T	Bromodichloromethane	0.562	0.607	-8.0	98	0.00
48 T	Methyl methacrylate	0.403	0.425	-5.5	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101323\
 Data File : VN079563.D
 Acq On : 13 Oct 2023 11:00
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 13 23:28:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.008	-14.3	87	0.00
50 S	Toluene-d8	1.300	1.384	-6.5	96	0.00
51 T	4-Methyl-2-Pentanone	0.531	0.550	-3.6	89	0.00
52 CM	Toluene	0.943	1.016	-7.7#	94	0.00
53 T	t-1,3-Dichloropropene	0.567	0.632	-11.5	96	0.00
54 T	cis-1,3-Dichloropropene	0.620	0.680	-9.7	95	0.00
55 T	1,1,2-Trichloroethane	0.378	0.399	-5.6	96	0.00
56 T	Ethyl methacrylate	0.520	0.603	-16.0	92	0.00
57 T	1,3-Dichloropropane	0.654	0.707	-8.1	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.199	0.247	-24.1	101	0.00
59 T	2-Hexanone	0.383	0.402	-5.0	86	0.00
60 T	Dibromochloromethane	0.380	0.422	-11.1	97	0.00
61 T	1,2-Dibromoethane	0.376	0.388	-3.2	94	0.00
62 S	4-Bromofluorobenzene	0.417	0.458	-9.8	97	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	92	0.00
64 T	Tetrachloroethene	0.344	0.337	2.0	93	0.00
65 PM	Chlorobenzene	1.114	1.151	-3.3	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.408	0.441	-8.1	100	0.00
67 C	Ethyl Benzene	1.935	2.139	-10.5#	95	0.00
68 T	m/p-Xylenes	0.703	0.805	-14.5	97	0.00
69 T	o-Xylene	0.692	0.768	-11.0	95	0.00
70 T	Styrene	1.092	1.280	-17.2	95	0.00
71 P	Bromoform	0.298	0.312	-4.7	96	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
73 T	Isopropylbenzene	4.652	4.454	4.3	95	0.00
74 T	N-amyl acetate	1.907	1.858	2.6	90	0.00
75 P	1,1,2,2-Tetrachloroethane	1.814	1.473	18.8	94	0.00
76 T	1,2,3-Trichloropropane	1.442	1.314	8.9	91	0.00
77 T	Bromobenzene	1.141	1.019	10.7	95	0.00
78 T	n-propylbenzene	5.263	5.329	-1.3	95	0.00
79 T	2-Chlorotoluene	3.368	3.168	5.9	97	0.00
80 T	1,3,5-Trimethylbenzene	3.668	3.657	0.3	94	0.00
81 T	trans-1,4-Dichloro-2-butene	0.497	0.448	9.9	89	0.00
82 T	4-Chlorotoluene	3.205	3.143	1.9	96	0.00
83 T	tert-Butylbenzene	3.005	2.911	3.1	91	0.00
84 T	1,2,4-Trimethylbenzene	3.501	3.679	-5.1	96	0.00
85 T	sec-Butylbenzene	4.192	4.194	-0.0	93	0.00
86 T	p-Isopropyltoluene	3.216	3.401	-5.8	94	0.00
87 T	1,3-Dichlorobenzene	1.809	1.838	-1.6	96	0.00
88 T	1,4-Dichlorobenzene	1.868	1.790	4.2	95	0.00
89 T	n-Butylbenzene	2.568	2.783	-8.4	92	0.00
90 T	Hexachloroethane	0.756	0.657	13.1	94	0.00
91 T	1,2-Dichlorobenzene	1.765	1.767	-0.1	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.269	0.232	13.8	85	0.00
93 T	1,2,4-Trichlorobenzene	0.547	0.592	-8.2	85	0.00
94 T	Hexachlorobutadiene	0.426	0.388	8.9	92	0.00
95 T	Naphthalene	1.669	1.608	3.7	77	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101323\
Data File : VN079563.D
Acq On : 13 Oct 2023 11:00
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Oct 13 23:28:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.652	0.633	2.9	84	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6