

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101419\
 Data File : VN058761.D
 Acq On : 14 Oct 2019 19:52
 Operator : JC/SP
 Sample : K5335-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IDW-GAC2-191010

Quant Time: Oct 15 04:19:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 12 00:00:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	386390	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	647345	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	588058	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	241809	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.01	65	297909	55.07	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	110.14%	
35) Dibromofluoromethane	7.57	113	210502	52.65	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	105.30%	
50) Toluene-d8	10.08	98	782169	48.92	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	97.84%	
62) 4-Bromofluorobenzene	12.40	95	260191	42.99	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	85.98%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.80	43	49557	22.129	ug/l	97
20) Methylene Chloride	4.53	84	35871	7.664	ug/l	99
27) cis-1,2-Dichloroethene	6.81	96	90658	18.730	ug/l	86
43) Isopropyl Acetate	8.15	43	18021	1.598	ug/l #	91
95) Naphthalene	15.13	128	46985	3.500	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101419\
Data File : VN058761.D
Acq On : 14 Oct 2019 19:52
Operator : JC/SP
Sample : K5335-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IDW-GAC2-191010

Quant Time: Oct 15 04:19:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 12 00:00:30 2019
Response via : Initial Calibration

