

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101419\
 Data File : VN058764.D
 Acq On : 14 Oct 2019 20:58
 Operator : JC/SP
 Sample : K5266-13RE
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AMENDED-COMPOSITE-ARE

Quant Time: Oct 15 04:24:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 12 00:00:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	389705	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	657362	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	593991	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	248965	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	301682	55.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.60%	
35) Dibromofluoromethane	7.64	113	295	0.07	ug/l	0.07
Spiked Amount 50.000			Recovery	=	0.14%	
50) Toluene-d8	10.08	98	783968	48.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.56%	
62) 4-Bromofluorobenzene	12.41	95	267142	43.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.94%	
Target Compounds						
16) Acetone	3.80	43	55887	24.743	ug/l	98
20) Methylene Chloride	4.53	84	79455	16.832	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101419\
Data File : VN058764.D
Acq On : 14 Oct 2019 20:58
Operator : JC/SP
Sample : K5266-13RE
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AMENDED-COMPOSITE-ARE

Quant Time: Oct 15 04:24:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 12 00:00:30 2019
Response via : Initial Calibration

