

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101419\
 Data File : VN058766.D
 Acq On : 14 Oct 2019 21:42
 Operator : JC/SP
 Sample : K5292-06RE
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20190855-AMENDED-COMP-BRE

Quant Time: Oct 15 04:28:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 12 00:00:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	382621	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	649637	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	589500	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	252077	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	303144	56.59	ug/l	0.00
Spiked Amount	50.000		Recovery	=	113.18%	
35) Dibromofluoromethane	7.65	113	731	0.18	ug/l	0.07
Spiked Amount	50.000		Recovery	=	0.36%	
50) Toluene-d8	10.08	98	788856	49.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.32%	
62) 4-Bromofluorobenzene	12.41	95	265744	43.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.50%	
Target Compounds						
16) Acetone	3.80	43	46093	20.785	ug/l	97
20) Methylene Chloride	4.53	84	7297	1.574	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101419\
Data File : VN058766.D
Acq On : 14 Oct 2019 21:42
Operator : JC/SP
Sample : K5292-06RE
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20190855-AMENDED-COMP-BRE

Quant Time: Oct 15 04:28:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 12 00:00:30 2019
Response via : Initial Calibration

