

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101420\
 Data File : VN064125.D
 Acq On : 14 Oct 2020 20:31
 Operator : JC/MD
 Sample : L4226-11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-08-101320

Quant Time: Oct 15 07:29:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101420W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 15 01:16:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	288209	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	546515	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	565517	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	209044	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	221570	53.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.96%	
35) Dibromofluoromethane	7.55	113	184119	48.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.34%	
50) Toluene-d8	10.06	98	692326	48.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.02%	
62) 4-Bromofluorobenzene	12.38	95	241496	49.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.46%	
Target Compounds						
20) Methylene Chloride	4.51	84	9626	2.555	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101420\
Data File : VN064125.D
Acq On : 14 Oct 2020 20:31
Operator : JC/MD
Sample : L4226-11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-08-101320

Quant Time: Oct 15 07:29:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101420W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 15 01:16:44 2020
Response via : Initial Calibration

