

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101420\
 Data File : VN064129.D
 Acq On : 14 Oct 2020 22:07
 Operator : JC/MD
 Sample : PB132343ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132343ZHE#01

Quant Time: Oct 15 07:37:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101420W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 15 01:16:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	298553	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	568497	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	626560	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	225794	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	232502	54.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.36%	
35) Dibromofluoromethane	7.55	113	194397	49.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.80%	
50) Toluene-d8	10.06	98	767488	51.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.38%	
62) 4-Bromofluorobenzene	12.38	95	270692	53.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.08%	
Target Compounds						
16) Acetone	3.79	43	11941	9.468	ug/l	97
20) Methylene Chloride	4.51	84	8410	2.155	ug/l	91
43) Isopropyl Acetate	8.13	43	91244	11.307	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101420\
Data File : VN064129.D
Acq On : 14 Oct 2020 22:07
Operator : JC/MD
Sample : PB132343ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132343ZHE#01

Quant Time: Oct 15 07:37:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101420W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 15 01:16:44 2020
Response via : Initial Calibration

