

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101420\
 Data File : VN064130.D
 Acq On : 14 Oct 2020 22:31
 Operator : JC/MD
 Sample : PB132343ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132343ZHE#02

Quant Time: Oct 15 07:42:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101420W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 15 01:16:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	273669	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	543318	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	593821	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	216179	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	218198	55.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.96%	
35) Dibromofluoromethane	7.55	113	183989	48.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.86%	
50) Toluene-d8	10.06	98	725820	51.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.30%	
62) 4-Bromofluorobenzene	12.38	95	254050	52.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.18%	
Target Compounds						
16) Acetone	3.78	43	11803	10.209	ug/l	92
18) Methyl Acetate	4.29	43	3703	1.277	ug/l #	86
20) Methylene Chloride	4.50	84	9141	2.555	ug/l #	83
43) Isopropyl Acetate	8.13	43	89795	11.643	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101420\
Data File : VN064130.D
Acq On : 14 Oct 2020 22:31
Operator : JC/MD
Sample : PB132343ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132343ZHE#02

Quant Time: Oct 15 07:42:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101420W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 15 01:16:44 2020
Response via : Initial Calibration

