


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN101420\
Data File : VN064111.D
Acq On : 14 Oct 2020 12:47
Operator : JC/MD
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101420W.M
Title : SW846 8260
Last Update : Thu Oct 15 01:16:44 2020



AutoFind: Scans 3307, 3308, 3309; Background Corrected with Scan 3293

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.9	16578	PASS
75	95	30	60	51.7	43162	PASS
95	95	100	100	100.0	83512	PASS
96	95	5	9	6.6	5479	PASS
173	174	0.00	2	1.4	920	PASS
174	95	50	100	78.1	65229	PASS
175	174	5	9	8.3	5442	PASS
176	174	95	101	95.9	62522	PASS
177	176	5	9	6.8	4242	PASS