

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101421\
 Data File : VN068922.D
 Acq On : 14 Oct 2021 14:10
 Operator : JC/MD
 Sample : M4192-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HAR-1

Quant Time: Oct 14 22:45:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	390823	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	633478	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	584771	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	230788	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	229723	44.582	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	89.160%
35) Dibromofluoromethane	8.024	113	188809	51.228	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.460%
50) Toluene-d8	10.443	98	772341	52.373	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.740%
62) 4-Bromofluorobenzene	12.733	95	266621	47.921	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.840%
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	19927	10.363	ug/l	95
20) Methylene Chloride	5.101	84	23680	5.350	ug/l	88
43) Isopropyl Acetate	8.563	43	39067	3.962	ug/l #	90
95) Naphthalene	15.509	128	6453	4.368	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101421\
Data File : VN068922.D
Acq On : 14 Oct 2021 14:10
Operator : JC/MD
Sample : M4192-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HAR-1

Quant Time: Oct 14 22:45:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

