

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101421\
 Data File : VN068923.D
 Acq On : 14 Oct 2021 14:35
 Operator : JC/MD
 Sample : M4192-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HAR-4

Quant Time: Oct 14 22:45:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	392456	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	639115	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	592426	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	235761	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	231592	44.757	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	89.520%		
35) Dibromofluoromethane	8.024	113	191060	51.382	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	102.760%		
50) Toluene-d8	10.443	98	779081	52.364	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	104.720%		
62) 4-Bromofluorobenzene	12.733	95	269212	47.960	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	95.920%		
Target Compounds						Qvalue
16) Acetone	4.293	43	21182	10.970	ug/l	99
20) Methylene Chloride	5.098	84	23925	5.383	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101421\
Data File : VN068923.D
Acq On : 14 Oct 2021 14:35
Operator : JC/MD
Sample : M4192-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HAR-4

Quant Time: Oct 14 22:45:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

